



宁波海曙鸿宇光电有限公司
NINGBO HAISHU HONYU OPTO-ELECTRO CO.,LTD.

J31.03.01 Manual Microtome (Metal)



For biological cut experiment, stable and reliable performance, easy and safe operating.

- It is made of Metal

- Diameter of working surface: $\phi 70\text{mm}$

Adjusted cut height

Minimum graduation: 0.01mm.

Maximum cut distance: 10mm.

-

-Packing box

-Cutting blades

